IJIRIS: International Journal of Innovative Research in Information Security
Volume 10, Issue 02, February 2024
E-ISSN: 2349-7017
P-ISSN: 2349-7009
https://www.ijiris.com/archives

Additive Manufacturing in Biomedical Application

Irfan S
PG Scholar, Dept of Medical Electronics,
Sengunthar Engineering College (Autonomous),
Tiruchengode, Tamil Nadu, INDIA
shajahanirfan2002@gmail.com
Dr.R.Sathish Kumar
Professor, Dept of Medical Electronics,
Sengunthar Engineering College (Autonomous),
Tiruchengode, Tamil Nadu, INDIA
harsiga@gmail.com
Prof.Reshma R
Professor, Dept of Medical Electronics,
Sengunthar Engineering College (Autonomous),
Tiruchengode, Tamil Nadu, INDIA
reshmapraveen027@gmail.com

Publication History
Manuscript Reference No: IJIRIS/RS/Vol.10/Issue02/FBIS10096
Research Article | Open Access | Double-Blind Peer-Reviewed | Article ID: IJIRIS/RS/Vol.10/Issue02/FBIS10096
Received: 24, January 2024|Revised:12, February 2024 | Accepted: 18, February 2024 Published Online: 29,February 2024
Volume 2024 | Article ID FBIS10096 http://www.ijiris.com/volumes/Vol10/iss-02/17.FBIS10096.pdf
Article Citation:Irfan,Sathish,Reshma(2024). Additive Manufacturing in Biomedical Application. International Journal of Innovative Research in Information Security,Volume 10, Issue 02, Pages 119-122
doi:> https://doi.org/10.26562/ijiris.2024.v1002.17
BibTex key: Irfan@2024Additive

Copyright: ©2024 This is an open access article distributed under the terms of the Creative Commons Attribution License; which Permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Abstract: Additive Manufacturing (AM), commonly known as 3D printing, has revolutionized the field of biomedical applications by offering innovative solutions for personalized and complex structures. This technology enables the fabrication of patient-specific implants, prosthetics, and tissues with enhanced precision and customization. The versatility of additive manufacturing allows the incorporation of biocompatible materials, fostering the development of implants tailored to individual anatomical requirements. Additionally, the rapid prototyping capabilities of 3D printing facilitate the creation of intricate models for surgical planning and education. The use of bioinks and biomaterials in additive manufacturing has paved the way for the fabrication of functional tissues and organs, advancing the prospects of regenerative medicine. Furthermore, the scalability and cost-effectiveness of 3D printing in the biomedical field hold significant promise for widespread adoption and accessibility. In conclusion, additive manufacturing stands as a transformative force in biomedical applications, offering unparalleled opportunities for personalized healthcare and advancing the frontiers of medical technology.

1. INTRODUCTION

Additive manufacturing, colloquially known as 3D printing, has emerged as a revolutionary technology at the intersection of engineering and medicine.



Fig. 1: A schematic flowchart of the study, outlining major contents and sub-contents to succinctly depict the work methodology.

In the biomedical domain, this transformative approach has significantly reshaped traditional manufacturing processes, offering unprecedented opportunities for customization and precision. By layering materials with meticulous accuracy, 3D printing enables the creation of intricate structures, ranging from patient-specific implants to artificial tissues. This paper delves into the multifaceted impact of additive manufacturing on biomedical applications, exploring its potential to redefine healthcare paradigms. The ability to seamlessly integrate diverse, biocompatible materials has paved the way for tailoring implants to individual anatomical nuances, promising enhanced patient outcomes [1-5]. Furthermore, the versatility of 3D printing extends beyond implant manufacturing, encompassing the production of prosthetics and facilitating advanced surgical planning through the generation of detailed anatomical models. The integration of bioinks and biomaterials in additive manufacturing has unlocked new frontiers in regenerative medicine, fostering the creation of functional tissues and organs. As we navigate through the following sections, we will unravel the diverse applications, challenges, and future prospects that underscore the pivotal role of additive manufacturing in shaping the landscape of biomedical innovation. Its scalability, cost-effectiveness, and potential for democratizing healthcare solutions make additive manufacturing a compelling force in the ongoing evolution of medical technology. Figure 2 offers a brief visual representation of the study's methodology, including major components and sub-components in a schematic flowchart.

2. LITERATURE SURVEY

Scaffolds, manufactured from high-molecular-weight PCL using the FDM 3D printing method, were designed for potential biomedical applications. [6] A series of experiments were conducted to assess biodegradation rates and biocompatibility. Furthermore, fragments of pure PCL were altered using low-temperature plasma to attain more hydrophilic properties, with surface wet ability evaluated through the sessile drop method. This [7] study seeks to investigate the parameters influencing the 3D bioprinting fabrication process via extrusion with a composite of Hydroxyapatite, gelatin, and gelatinmethacryloyl (GelMa) scaffold. Additionally, it aims to evaluate biocompatibility with osteoblast and fibroblast cells for bone tissue engineering. Preliminary results involve the preparation of gelatin with Hydroxyapatite, and the concentration of Hydroxyapatite and gelatinbioink was determined. The paper [8] reviews the additive manufacturing method known as laser cladding, utilized for surface modification of metallic tools, especially in applications requiring abrasive and corrosive resistance. Despite titanium's favorable properties, its limitations in tribology-related industries necessitate ongoing research for enhancing the Ti6Al4V alloy. Bioprinting, adopted to address ethical concerns and enhance the automation, reproducibility, and geometry stability of traditional in vitro models, faces challenges related to heterogeneity, microstructural complexity, and integration. Overcoming these challenges is crucial for advancing the technological readiness of [9] bioprinting and generating multi-organ platforms to understand biological responses. The [10] development of sustainable sensor solutions is emphasized, presenting a Chitosan-based sensor for biomedical applications. Fabricated on a bio-sourced and biocompatible thermoplastic polyurethane substrate, this sensor's response to Acetone vapor is studied. The use of Chitosan and additive manufacturing technologies renders the sensor sustainable and well-suited for biomedical applications due to the incorporation of biocompatible materials.

3. AM IN HIP PROSTHESIS

The application of AM, or 3D printing, in the production of hip prostheses represents a significant advancement in orthopedic surgery. Traditional methods of manufacturing hip implants often face limitations in terms of customization and precision. AM technologies offer a transformative solution by allowing for the creation of patient-specific hip prostheses tailored to individual anatomical variations. The utilization of AM in hip prosthesis manufacturing involves the layer-by-layer deposition of biocompatible materials, such as titanium or cobalt-chromium alloys. This enables the fabrication of complex geometries that mimic the natural hip joint, ensuring a more precise fit and improved functionality. Customization extends beyond size, encompassing variations in bone density, shape, and alignment, thereby enhancing the overall success and longevity of the prosthetic implant. Moreover, additive manufacturing facilitates the incorporation of porous structures in the implant design. Porous surfaces encourage bone in growth, promoting osseo integration and long-term stability. This innovation reduces the risk of implant loosening and enhances the patient's quality of life by mimicking the natural biomechanics of the hip joint. As we delve into the intricate landscape of additive manufacturing in hip prosthesis, this paper will explore the technological advancements, clinical outcomes, and challenges associated with integrating 3D printing into orthopedic practice. The customizable nature, improved patient outcomes, and the potential for accelerated manufacturing processes position additive manufacturing as a promising avenue in the evolution of hip prosthesis design and implementation.

4. PROPOSED METHODOLOGY

The working process of additive manufacturing in hip prosthesis involves several key steps, showcasing the transformative capabilities of 3D printing technology in orthopedic applications:

Patient Imaging and Data Acquisition: The process begins with the acquisition of detailed imaging data, typically from sources like CT scans or MRI scans of the patient's hip joint. These scans provide precise information about the anatomy, including bone structure, density, and any abnormalities.

Digital Design and Modeling: Using specialized software, orthopedic engineers and surgeons collaborate to design a customized hip prosthesis based on the patient's specific anatomy. This digital model serves as the blueprint for the additive manufacturing process.

Material Selection: Biocompatible materials, such as titanium or cobalt-chromium alloys, are chosen for the additive manufacturing process. These materials ensure compatibility with the human body and provide the required strength and durability for the hip prosthesis.

Layer-by-Layer Printing: The 3D printing machine employs layer-by-layer deposition techniques to fabricate the hip prosthesis. This additive process allows for the creation of intricate geometries and precise structures, ensuring a customized fit for the patient.

Post-Processing and Finishing: Once the additive manufacturing process is complete, the printed hip prosthesis undergoes post-processing steps. This may involve removing support structures, smoothing surfaces, and applying any necessary finishes to optimize the prosthetic for implantation.

Quality Control and Testing: Rigorous quality control measures are implemented to ensure the final product meets the required standards. This may involve various tests, such as dimensional accuracy checks, material integrity assessments, and verification of the prosthesis against the original digital model.

Implantation: The custom-designed and 3D-printed hip prosthesis is then ready for surgical implantation. The surgeon follows the preoperative plan based on the digital model, ensuring precise placement and alignment within the patient's hip joint.

Follow-Up and Monitoring: Post-surgery, the patient undergoes regular follow-up examinations and monitoring to assess the functionality of the implanted hip prosthesis. The use of 3D printing allows for a more personalized fit, potentially contributing to improved patient outcomes and reduced postoperative complications. This comprehensive working process underscores the advantages of additive manufacturing in hip prosthesis, offering a patient-centric approach that aligns with individual anatomical variations and improves the overall success of the orthopedic intervention.

5. RESULT AND DISCUSSIONS

The integration of additive manufacturing (AM) in the production of hip prostheses has yielded promising outcomes and sparked significant discussions within the orthopedic community. The following section delves into the results observed in clinical applications and opens avenues for discussions on the implications, challenges, and future directions in this evolving field, Shown in figure 2.

1. Customization and Patient-Specific Fit: Results indicate a substantial improvement in the customization and patient-specific fit of hip prostheses through AM. The ability to tailor the design based on individual anatomical data has shown to enhance implant accuracy, reducing issues related to mismatch, discomfort, and potential complications.

2. Osseointegration and Biomechanical Performance: Studies highlight positive outcomes in terms of osseointegration, with porous structures incorporated into 3D-printed implants promoting bone in growth. This feature contributes to improved biomechanical performance, reducing the risk of implant loosening and enhancing long-term stability.

3. Accelerated Manufacturing and Reduced Lead Times: The additive manufacturing process has demonstrated its efficiency in terms of accelerated manufacturing and reduced lead times. This is particularly beneficial in urgent cases or scenarios requiring rapid production, allowing for quicker surgical interventions and patient recovery.

4. Challenges and Considerations: Discussions surrounding the adoption of AM in hip prostheses acknowledge challenges such as material selection, regulatory considerations, and the need for standardized testing protocols. The biocompatibility, mechanical properties, and long-term durability of 3D-printed materials are subjects of ongoing research and discussion.

5. Cost Implications and Scalability: While the cost-effectiveness of AM in hip prostheses is a topic of interest, initial investments in technology and materials may pose challenges. Discussions emphasize the potential long-term cost benefits and explore avenues for improving scalability to make 3D-printed prosthetics more economically viable.



Fig. 2 depicts diverse applications of additively manufactured Ti-based alloys, outlining challenges and suggesting solutions tailored for load-bearing uses.

6. CONCLUSION

In conclusion, additive manufacturing has emerged as a groundbreaking force in the realm of hip prostheses, revolutionizing customization and precision in implant design. The observed improvements in patient-specific fits, osseointegration, and accelerated manufacturing highlight the transformative impact of 3D printing on orthopedic practices. However, ongoing discussions emphasize the need for addressing challenges such as material selection, regulatory considerations, and standardization. The future enhancement of additive manufacturing in hip prostheses lies in further advancements in material sciences, refinement of printing techniques, and the establishment of comprehensive testing protocols. As research progresses, the exploration of innovative design methodologies and the conduct of long-term clinical studies will contribute to solidifying the technology's role in enhancing patient outcomes and shaping the future of orthopedic interventions.

REFERENCES

[1]. D. U. Ozsahin, E. P. Onakpojeruo, B. B. Duwa, B. Uzun, I. Ozsahin and E. C. Chioma, "Comparative Evaluation of 3D Filaments Used in Additive Manufacturing of Biomedical Tools; Using Fuzzy PROMETHEE," 2023 Advances in Science and Engineering Technology International Conferences (ASET), Dubai, United Arab Emirates, 2023, pp. 1-7, https://doi:10.1109/ASET56582.2023.10180544.

[2]. R. Vraitch, M. Prince and J. -B. R. G. Souppez, "Infill Density in Additive Manufacturing and Application to the DFMA of an Iron Man Helmet," 2023 28th International Conference on Automation and Computing (ICAC), Birmingham, United Kingdom, 2023, pp. 1-6, https://doi:10.1109/ICAC57885.2023.10275307.

[3]. H. Marangalou, M. Arturo Gonzalez and U. Guler, "Additively Manufactured Receiver Design for Wirelessly-Powered Biomedical Applications," 2023 IEEE Biomedical Circuits and Systems Conference (BioCAS), Toronto, ON, Canada, 2023, pp. 1-5, https://doi:10.1109/BioCAS58349.2023.10388694.

[4]. Massaroni, L. Vitali, D. L. Presti, S. Silvestri and E. Schena, "Design, Development and Characterization of a Novel Fully Additively Manufactured Deformable Conductive Force Sensor," 2023 International Workshop on Biomedical Applications, Technologies and Sensors (BATS), Catanzaro, Italy, 2023, pp. 22-27, https://doi:10.1109/BATS59463.2023.10303121

[5]. L. Liu, E. Jo, U. Vaidya, S. Kim and F. Ju, "An Integrated Offline and Online Optimization Framework for Large Scale Additive Manufacturing," 2023 IEEE 19th International Conference on Automation Science and Engineering (CASE), Auckland, New Zealand, 2023, pp. 1-6, https://doi:10.1109/CASE56687.2023.10260301

[6]. Y. Kravchenko, M. Pogorielov, A. Taran, M. Kubakh, I. Tepliakova and Y. Husak, "Additive Manufacturing of Bioresorbable Scaffolds Based on Polycaprolactone and Composites," 2023 IEEE 13th International Conference Nanomaterials: Applications & Properties (NAP), Bratislava, Slovakia, 2023, pp. IMT02-1-IMT02-4, https://doi:10.1109/NAP59739.2023.10311024

[7]. K. Tappaviboon et al., "Validation of parameter for bioink Hydroxyapatite/Gelatin Scaffold by 3D extrusion bio-printing for orthopedic surgery," 2023 15th Biomedical Engineering International Conference (BMEiCON), Tokyo, Japan, 2023, pp. 1-4, https://doi:10.1109/BMEiCON60347.2023.10322099.

[8]. S. B. Boshoman, O. S. Fatoba and T. -C. Jen, "Laser Cladding of Titanium Alloys: A Review," 2023 14th International Conference on Mechanical and Intelligent Manufacturing Technologies (ICMIMT), Cape Town, South Africa, 2023, pp. 158-163, https://doi:10.1109/ICMIMT59138.2023.10200605

[9]. J. Caceres-Alban, M. Sanchez and F. L. Casado, "Bioprinting: A Strategy to Build Informative Models of Exposure and Disease," in IEEE Reviews in Biomedical Engineering, vol. 16, pp. 594-610, 2023, https://doi:10.1109/RBME.2022.3146293.

[10]. J. Zikulnig, S. Lengger, L. Rauter, S. Carrara and J. Kosel, "A Sustainable Printed Chitosan-Based Sensor for Acetone Detection," 2023 18th Conference on Ph.D Research in Microelectronics and Electronics (PRIME), Valencia, Spain, 2023, pp. 121-124, https://doi:10.1109/PRIME58259.2023.10161879.